

# **Boston ADVANCE®**

## **Comfort Formula Conditioning Solution**

For Fluoro Silicone Acrylate and Silicone Acrylate  
Rigid Gas Permeable Contact Lenses

**IMPORTANT: Please read carefully and keep this information for future use.**

### **Description:**

A sterile, aqueous, buffered, slightly hypertonic solution containing a cationic cellulose derivative polymer, a cellulosic viscosifier, polyvinyl alcohol and a derivatized polyethylene glycol as wetting and cushioning agents; preserved with chlorhexidine gluconate (0.003%), polyaminopropyl biguanide (0.0005%) and edetate disodium (0.05%).

### **Actions:**

This solution contains a patented multi-polymer cushioning system and a patented optimized preservative system to minimize ocular irritation while being effective in destroying harmful organisms when used with Boston ADVANCE® Cleaner.

### **Indications:**

For wetting, disinfecting and soaking (after cleaning and rinsing) fluoro silicone acrylate and silicone acrylate rigid gas permeable contact lenses.

### **Contraindications:**

Do not use this solution if you are allergic to any ingredient in this product. Not For Use With Soft (Hydrophilic) Contact Lenses.

### **Warnings:**

**To avoid contamination, never touch dropper tip of the container with your hands or to any surface.** Keep bottle tightly closed when not in use.

Contact lenses and solutions, if not handled properly, can become contaminated by common bacteria known to cause ocular infections. Serious damage to the eye and loss of vision may result from using contaminated solutions or from wearing contaminated or damaged contact lenses.

**PROBLEMS WITH CONTACT LENSES AND LENS CARE PRODUCTS COULD RESULT IN SERIOUS INJURY TO YOUR EYES. It is essential that you follow your eye care professional's directions and all labeling instructions for proper use of your lenses and lens care products.**

**EYE PROBLEMS, INCLUDING CORNEAL ULCERS CAN DEVELOP RAPIDLY AND LEAD TO LOSS OF VISION. THEREFORE, IF YOU EXPERIENCE EYE DISCOMFORT, EXCESSIVE TEARING, VISION CHANGES, OR REDNESS OF THE EYES, IMMEDIATELY REMOVE YOUR LENSES AND PROMPTLY CONTACT YOUR EYE CARE PRACTITIONER.**

### **Precautions:**

To ensure proper hygiene and handling of contact lenses and contact lens solutions, always follow label directions carefully.

- Never reuse this solution.

- Always use fresh solution for soaking and storing lenses.
- Always wash your hands before handling your lenses.
- Store solutions at room temperature. Avoid freezing.
- After inserting your lenses, always thoroughly clean the interior of the lens case with hot water and air dry after each use.
- Use before expiration date on the container.
- **NEVER SOAK LENSES IN ANY BOSTON® CLEANER.**
- **ALWAYS RINSE LENSES THOROUGHLY WITH FRESH TAP WATER AFTER CLEANING.**
- **KEEP OUT OF REACH OF CHILDREN.**

### **Adverse Effects:**

**The following problems may occur:**

- Eyes stinging, burning, itching (irritation)
- Excessive watering (tearing) of the eyes
- Unusual eye secretions
- Redness of the eyes
- Reduced sharpness of vision (visual acuity)
- Blurred vision
- Sensitivity to light (photophobia)
- Dry eyes

If you notice any of the above conditions, immediately remove and examine the lenses. If the problem stops and the lenses appear to be undamaged, thoroughly clean, rinse and disinfect the lenses and reinsert. If the problem continues or if a lens appears to be damaged, **DO NOT REINSERT.** Contact your eye care professional immediately. If any of the above problems occur, a serious condition such as infection, corneal ulcer, neovascularization or iritis may be present. Seek immediate professional identification of the problem and treatment to avoid serious eye damage.

### **DIRECTIONS FOR USE**

**To ensure proper disinfecting, all steps listed below must be followed:**

1. Wash your hands with mild soap. (Caution - pump soaps may contain oil-based suspension agents.)
2. After cleaning with Boston ADVANCE Cleaner and rinsing, place lenses in empty lens case and fill to top of ridges with fresh Boston ADVANCE Comfort Formula Conditioning Solution. **Soak lenses for at least four (4) hours (or overnight) before wearing. Always use fresh solution for soaking and storing lenses.**
3. After removing lenses from the lens case, apply fresh Boston ADVANCE Comfort Formula Conditioning Solution to wet the lenses for additional cushioning, if desired, and insert.
4. Thoroughly clean interior of lens case with hot water and air dry after each use.
5. Discard solution ninety (90) days after opening. Record date opened in space provided on bottle label. Note: When used daily, this solution will be depleted before ninety (90) days.

- **For weekly protein removal, use Boston® ONE STEP Liquid Enzymatic Cleaner.**
- **Lens Case should be replaced frequently to avoid potential microbial contamination and ocular irritation.**

**How Supplied:**

Boston ADVANCE Comfort Formula Conditioning Solution is supplied sterile in 4 fluid ounce (120 mL), 3 fluid ounce (90 mL) and 1 fluid ounce (30 mL) bottles. All retail Boston® Solutions are packed in unit cartons. Bottles are sealed with an imprinted neckband and marked with a lot number and expiration date.

**Do not use if imprinted neckband on bottle is missing or broken.**

© Polymer Technology  
Made in USA 5/00  
Manufactured for:  
Polymer Technology  
Rochester, NY 14603-0450.

Questions or Comments:  
1-800-333-4730  
email: [www.bausch.com](http://www.bausch.com)  
3414506W

